



SEAWARD
ELECTRICAL SAFETY TESTING
& MEASURING.

How do I PAT test power tool chargers and mobile phone chargers?

Class 1 chargers should have continuity to the DC plug, so clip your earth lead there. Class 2, again clip your earth lead there as this may well be the only exposed metal part. Should the charger contain a transformer, then a thorough visual inspection will suffice.

If you require more help, please contact us at <https://www.seaward.com/gb/enquiry/>.